



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/696,532	10/28/2003	Naobumi Okada	03653/LH	6456
1933 7590 06/27/2007 FRISHAUF, HOLTZ, GOODMAN & CHICK, PC 220 Fifth Avenue 16TH Floor NEW YORK, NY 10001-7708			EXAMINER ELVE, MARIA ALEXANDRA	
			ART UNIT 1725	PAPER NUMBER
			MAIL DATE 06/27/2007	DELIVERY MODE PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No.		Applicant(s)	
	10/696,532		OKADA, NAOBUMI	
	Examiner		Art Unit	
	M. Alexandra Elve		1725	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☐ Responsive to communication(s) filed on ____.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-36 is/are pending in the application.
- 4a) Of the above claim(s) ____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) ____ is/are allowed.
- 6) ☒ Claim(s) 1-36 is/are rejected.
- 7) ☐ Claim(s) ____ is/are objected to.
- 8) ☐ Claim(s) ____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 28 October 2003 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☒ All b) ☐ Some * c) ☐ None of:
1. ☒ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. ____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. ____. |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date ____. | 6) <input type="checkbox"/> Other: ____. |

DETAILED ACTION

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 1-5, 7, 10-16, 18, 20-36 are rejected under 35 U.S.C. 103(a) as being unpatentable over Bonner et al. (USPN 6,251,516) in view of Baer et al. (USPN 6,870,625).

Bonner et al. disclose isolation of cellular material under microscopic visualization. An imaging system is used to receive sample images. A laser is used to cut the desired section, which is then transfer to a slide for further examination and analysis. The system allows unaltered microscopic observation prior to, during and following processing. Additionally, microscopic objectives are used.

Although Bonner et al. disclose the use of microscopic objectives; specific objective lenses are not taught.

Baer et al. discloses an automate laser capture microdissection system. The entire tissue sample or specific parts of it can be selectively excited by selecting different lenses of one or more objectives from an objective turret wheel (268).

Art Unit: 1725

It would have been obvious to one of ordinary skill in the art at the time of the invention to use an objective lens(es) as taught by Baer et al. in the Bonner et al. system because it is merely a type of microscopic objective assembly.

Intended use has been continuously held not to be germane to determining the patentability of the apparatus, In re Finsterwalder, 168 USPQ 530.

The manner or method in which a machine is to be utilized is not germane to the issue of patentability of the machine itself, In re Casey, 152 USPQ 235, 238.

Purpose to which an apparatus is to be put and expression relating apparatus to contents thereof during intended operation are not significant in determining patentability of an apparatus claim, *Ex parte Thibault*, 164 USPQ 666.

A recitation with respect to the manner in which a claimed apparatus is intended to be employed does not differentiate the claimed apparatus from a prior art apparatus satisfying the claimed structural limitations, *Ex parte Masham*, 2 USPQ 2d 1647.

Claims 6, 8-9, 17, 19 are rejected under 35 U.S.C. 103(a) as being unpatentable over Bonner et al. and Baer et al., as stated above and further in view of Caprioli (USPN 6,756,586).

Bonner et al. does not teach the evaporation of the imaging sample. Caprioli discloses the use of a laser beam to release samples for analysis. It would have been obvious to one of ordinary skill in the art at the time of the invention to use the release technique as taught by Caprioli in the Bonner et al. system because this yields more data for study and evaluation.

Art Unit: 1725

Intended use has been continuously held not to be germane to determining the patentability of the apparatus, In re Finsterwalder, 168 USPQ 530.

The manner or method in which a machine is to be utilized is not germane to the issue of patentability of the machine itself, In re Casey, 152 USPQ 235, 238.

Purpose to which an apparatus is to be put and expression relating apparatus to contents thereof during intended operation are not significant in determining patentability of an apparatus claim, *Ex parte Thibault*, 164 USPQ 666.

A recitation with respect to the manner in which a claimed apparatus is intended to be employed does not differentiate the claimed apparatus from a prior art apparatus satisfying the claimed structural limitations, *Ex parte Masham*, 2 USPQ 2d 1647.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to M. Alexandra Elve whose telephone number is 571-272-1173. The examiner can normally be reached on 6:30-3:00 Monday to Friday.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Jonathan Johnson can be reached on 571-272-1177. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 1725

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

June 25, 2007.



M. Alexandra Elve
Primary Examiner 1725